

<b>CERTIFICATE OF ANALYSIS</b>	<b>GLUCOSE MR</b>
--------------------------------	-------------------

<b>REF</b>	1129005	2 x 50 mL	X	1129088	1 x 20 L	X	KR10202	8 x 30 mL	X	1929005	1 x 3mL	
	1129010	4 x 100 mL	X	CT10200	2 x 40 mL	X	KR10205	18 x 30 mL	X	1929080	1 x 500mL	
	1129015	4 x 250 mL	X	CT10202	6 x 40 mL	X				1929081	1 x 1L	
	1129085	1 x 5 L	X	CT10205	10 x 60 mL	X				1929085	1x 5L	
	1129050	50 ml/80 un.	x	1129055	100 ml/40 un.	x	1129060	250 ml/28 un.	x	1929006	3 ml/60 un	

<b>Cont</b>	<b>R1</b>	<b>X</b>	<b>LOT</b>	15829		2020-04	<b>IVD</b>
	<b>Standard</b>	<b>X</b>	<b>LOT</b>				2-8°C

**PHYSICAL AND CHEMICAL CHARACTERISTICS**

TEST		SPECIFICATIONS	RESULT	
Appearance	<b>R1</b>	Liquid. Clear. Free of particles. Clear to faintly pink.	ACCEPTED	X
	<b>CAL</b>	Liquid. Clear. Free of particles.	ACCEPTED	X

(ACCEPTED / N/A)

**FUNCTIONALITY**

<p><b>1. Analytical parameters</b></p> <p><i>Blank.</i> Absorbance of Working reagent against water.</p> <p><i>Sensitivity.</i> A<sub>500</sub> / 100 mg/dL Glucose<sub>Standard</sub></p> <p><i>Linearity.</i> Higher level established by the analytical performance.</p> <p><i>Accuracy.</i> Ratio between assigned and obtained Glucose values.</p> <p><b>2. Stability Testing</b></p> <p><i>Fresh Reagents (FR).</i> Items kept at 2-8°C.</p> <p><i>Stressed Reagents (SR).</i> Items kept at 37/42°C for a stated period of time.</p>
---

<b>INSTRUMENT</b>	SHIMADZU / LIDA BS-300	500 nm against reagent blank
-------------------	------------------------	------------------------------

TEST	SPECIFICATIONS	RESULTS	
		FR	SR
Blank	A <sub>500</sub> <0.100	ACCEPTED	ACCEPTED
Sensitivity	0.280-0.400	ACCEPTED	ACCEPTED
Linearity	Up to 500 mg/dL	ACCEPTED	ACCEPTED
Accuracy	Within the expected values	ACCEPTED	ACCEPTED

(ACCEPTED / N/A)

This lot was assayed by the test method described in the manufacturer's package insert. These results are applicable to this lot and subsequent sublots, for which a capital letter has been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

<b>This lot meets our specifications and is released for sale</b>	<b>APPROVED</b>	<p><b>Signed</b>  <b>Date of Manufacture</b>                  Francesc Guerrero                  2018-04-09                  Quality Control Dept.</p>
---	-----------------	--

