

<b>CERTIFICATE OF ANALYSIS</b>	<b>TRIGLYCERIDES</b>
--------------------------------	----------------------

<b>REF</b>	1155005	2 x 50 mL	X	CT10360	2 X 40 mL	X	KR10362	8 x 30 mL	X	1955005	1 x 3 mL	X
	1155010	4 x 100 mL	X	CT10362	6 x 40 mL	X	KR10365	18 x 30 mL	X	1955080	1 x 0.5 L	x
	1155085	1 x 5 L	X	CT10365	10 x 60 mL	X	1155060	250 ml/28 un	x	1955081	1 x 1 L	x
	1155088	1 x 20 L	X	1155050	50 ml/80 un	x	1155055	100 ml/40 un	x	1955006	3 ml/60 un.	x

<b>Cont</b>	<b>R1</b>	<b>X</b>	<b>LOT</b>	15820		2020-04	<b>IVD</b>		2-8°C
	<b>Standard</b>	<b>X</b>	<b>LOT</b>						

**PHYSICAL AND CHEMICAL CHARACTERISTICS**

TEST		SPECIFICATIONS	RESULT	
Appearance	<b>R1</b>	Slightly yellow. Liquid. Clear. Free of particles.	Clear	ACCEPTED
	<b>CAL</b>	Liquid. Clear. Free of particles.	Clear	ACCEPTED

(ACCEPTED / N/A)

**FUNCTIONALITY**

- 1. Analytical parameters**  
*Blank.* Absorbance of the working reagent against water.  
*Sensitivity.* Absorbance of the Triglycerides Standard 200 mg/dL against the blank.  
*Linearity.* Absorbance of the Internal Standard 800 mg/dL.  
*Accuracy.* Ratio between assigned and obtained Triglycerides values.
- 2. Stability Testing**  
*Fresh Reagents (FR).* Items kept at 2-8°C.  
*Stressed Reagents (SR).* Items kept at 37°C for a stated period of time.

<b>INSTRUMENT</b>	SHIMADZU / BS300 / SK500	510 nm against reagent blank
-------------------	--------------------------	------------------------------

TEST	SPECIFICATIONS	RESULTS			
		FR		SR	
Blank	0.020 - 0.120	0.080	ACCEPTED	0.090	ACCEPTED
Sensitivity	0.240 - 0.325	0.315	ACCEPTED	0.305	ACCEPTED
Linearity	Up to 800 mg/dL	802	ACCEPTED	805	ACCEPTED
Accuracy	Within the expected values [79.6 – 110 ] mg/dL	92	ACCEPTED	97	ACCEPTED

(ACCEPTED / N/A)

This lot was assayed by the test method described in the manufacturer's package insert.

These results are applicable to this lot and subsequent sublots, for which a capital letter has been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

<b>This lot meets our specifications and is released for sale</b>	APPROVED	<b>Signed</b> <b>Date of Manufacture</b> Francesc Guerrero 2018-03-27 Quality Control Dept.
---	----------	---

