

<b>CERTIFICATE OF ANALYSIS</b>	<b>S. PARATYPHI B-H</b>
--------------------------------	-------------------------

<b>REF</b>	2119005	5 mL	<b>X</b>	2119081	<b>X</b>
	2120005	50 mL		2119082	<b>X</b>
				2119085	<b>X</b>

<b>LOT</b> 21882	 2020-09	<b>IVD</b>
 2-8°C		

**PHYSICAL AND CHEMICAL CHARACTERISTICS**

REAGENT	TEST	SPECIFICATIONS	RESULTS
S. PARATYPHI B-H	APPEARANCE	Homogeneous suspension, free of macroscopic particles.	ACCEPTED
	COLOUR	Red	
	pH	8.2 +/- 0.2	

(ACCEPTED / N/A)

**FUNCTIONALITY**

REAGENT	TEST: SLIDE AND TUBE METHOD	SPECIFICATIONS	RESULTS
S. PARATYPHI B-H	Sensitivity, against Internal Reference Material	Titer: Within tolerance range	ACCEPTED
	Positive Control	3+	
	Negative Control	Negative	

(ACCEPTED / N/A)

**STABILITY TEST** (14 days at 37°C)

REAGENT	TOLERANCE	RESULT
S. PARATYPHI B-H	After a stability study, the specifications remain unchanged	ACCEPTED

(ACCEPTED / N/A)

This lot was tested by test methods described in the manufacturer's package insert.

These results are applicable to this lot and sequential sublots, for which a capital letter have been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

<b>This lot meets our specifications and is released for sale</b>	APPROVED	<b>Signed</b> <b>Date of Manufacture</b> Pilar Montolio 2018-02-27 Quality Control Dept.
	(APPROVED / NOT APPROVED)	

