

<b>CERTIFICATE OF ANALYSIS</b>	<b>Febrile Antigens</b>
--------------------------------	-------------------------

<b>REF</b>	2199435	6 vials x 5 mL	X
	2199610	6 vials x 5 mL	X

**LOT** 21861  
 2-8°C

2019-10

**IVD**

**PHYSICAL AND CHEMICAL CHARACYTEERISTICS**

REAGENT	TEST	SPECIFICATIONS	RESULTS
Febrile Antigens	APPEARANCE	Homogeneous suspension, free of macroscopic particles.	ACCEPTED
	COLOUR	Blue/Red	
	pH	Inside tolerance range	

(ACCEPTED / N/A)

**FUNCTIONALITY**

REAGENT	TEST: SLIDE AND TUBE METHOD	SPECIFICATIONS	RESULTS
Febrile Antigens	Sensitivity, against Internal Reference Material	Titer: Inside tolerance range	ACCEPTED
	Positive Control	3+	
	Negative Control	Negative	

(ACCEPTED / N/A)

**STABILITY TEST** (14 days at 37°C)

REAGENT	TOLERANCE	RESULT
Febrile Antigens	After a stability study, the specifications remain unchanged	ACCEPTED

(ACCEPTED / N/A)

This lot was tested by test methods described in the manufacturer's package insert.

These results are applicable to this lot and sequential sublots, for which a capital letter have been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

<b>This lot meets our specifications and is released for sale</b>	APPROVED	<b>Signed - Date of Manufacture</b> T.Simó 17 March 2017 Quality Control Dept.
	(APPROVED / NOT APPROVED)	

