

CERTIFICATE OF ANALYSIS	IgA + IgG + IgM id
--------------------------------	---------------------------

REF	3300605	3 x 5 Test	x				
-----	---------	------------	---	--	--	--	--

LOT G322.16
2-8°C

2018-12
IVD

FUNCTIONALITY

	TEST	SPECIFICATIONS	RESULTS
IgA	Slope	--	0.118336652
	Interp. Coeff. r^2	0.97 - 1	0.99239076
	Interception	10.0 – 11.0	10.0457647

	TEST	SPECIFICATIONS	RESULTS
IgG	Slope	--	0.034766381
	Interp. Coeff. r^2	0.97 - 1	0.99334554
	Interception	10.0 – 11.0	10.4912941

	TEST	SPECIFICATIONS	RESULTS
IgM	Slope	--	0.164091619
	Interp. Coeff. r^2	0.97 - 1	0.99176449
	Interception	10.0 – 11.0	10.0689412

This lot was assayed by the test method described in the manufacturer's package insert.

These results are applicable to this lot and subsequent sublots, for which a capital letter has been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

<p>This lot meets our specifications and is released for sale</p>	<p>APPROVED</p>	<p>Signed Date of Manufacture, Mercè Oriol 12 January 2017 Quality Control</p>
	<p>(APPROVED / NOT APPROVED)</p>	



QUALITY SYSTEM CERTIFIED
ISO 9001 ISO 13485



LINEAR CHEMICALS, S.L.U. Joaquim Costa 18 2ª planta. 08390 Montgat (Barcelona) SPAIN
Telf. (+34) 934 694 990; E-mail: info@linear.es ; website: www.linear.es