



<b>CERTIFICATE OF ANALYSIS</b>	<b>ANTITROMBINA III id</b>
--------------------------------	----------------------------

<b>REF</b>	3307005	1 x 15 test			

**LOT** F279.16  
 2-8°C

 2018.10  
**IVD**

**FUNCTIONALITY**

REAGENT	TEST	SPECIFICATIONS	RESULTS
<b>ATT III id</b>	Slope	--	1.61610505055
	Interp. Coeff. r <sup>2</sup>	0.97-1.00	0.99448385
	Interception	10.0 – 11.0	10.5255087

This lot was tested by test methods described in the manufacturer's package insert.  
 These results are applicable to this lot and sequential sublots, for which a capital letter have been assigned after the lot number.  
 This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

This lot meets our specifications and is released for sale	<b>APPROVED</b>	Signed - Date of Manufacture Mercè Oriol Quality Control Department 12 January 2017
	(APPROVED / NOT APPROVED)	

