

<b>CERTIFICATE OF ANALYSIS</b>	<b>STEL 3-Cleaner 1 L</b>
--------------------------------	---------------------------

<b>REF</b>	3802015	1 L	X
------------	---------	-----	---

<b>Cont</b>	<b>CLEANER</b>	<b>X</b>	<b>LOT</b>	38033		2020/04	<b>IVD</b>
-------------	----------------	----------	------------	-------	---	---------	------------

**PHYSICAL AND CHEMICAL CHARACTERISTICS**

TEST		SPECIFICATIONS	RESULT	
Appearance	<b>DIL</b>	Transparent liquid; No precipitation, particles, or floccule	ACCEPTED	X
pH	<b>DIL</b>	8.5 ± 0.5	ACCEPTED	X

(ACCEPTED / N/A)

**FUNCTIONALITY**

<b>INSTRUMENT</b>	<b>STEL 3</b>
-------------------	---------------

TEST	SPECIFICATIONS	RESULTS
Blank counting	Result of a measurement using Haematology Analyser, test value should be: White blood cell count(WBC)≤0.3*10 <sup>9</sup> /L, Red blood cell count(RBC) ≤0.05*10 <sup>12</sup> /L, Platelet count (PLT) ≤10*10 <sup>9</sup> /L, Hemoglobinometry (HGB) ≤2g/L.	WBC:0 0 0 RBC:0 0 0 PLT:0 0 0 HGB:0 0 0
Component	Buffer solution <0.3%,Protease <0.2%,	pass

(ACCEPTED / N/A)

This lot was assayed by the test method described in the manufacturer's package insert. These results are applicable to this lot and subsequent sublots, for which a capital letter has been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

<b>This lot meets our specifications and is released for sale</b>	APPROVED	<b>Signed</b> <b>Date of Manufacture</b> Mercè Oriol Quality Control Dept. 02 May 2017
---	----------	--

