

CERTIFICATE OF ANALYSIS


3802025

STEL 3-Blood Control

3x3 mL

LOT

38042


2018-06

IVD
QUALITATIVE CONTROLS

TEST	SPECIFICATIONS	RESULTS			
		PASS	x	FAIL	
APPEARANCE	Liquid Red				

<i>MAIN COMPONENTS</i>	Human erythrocytes, leukocyte analogs, platelet analogs and preservatives.
<i>HUMAN SOURCE</i>	HBs Ag, Anti-HIV 1/2 and Anti HCV Nonreactive when units tested individually.
<i>STORAGE TEMPERATURE</i>	Refrigerate 2-8°C.
<i>STABILITY</i>	The expiry date is illustrated on the side of each pack.
<i>SAFETY</i>	Because no test method can offer complete assurance that products derived from human source will not transmit infectious agents, it is recommended that this product be handled with the same precautions used for patients specimens.
<i>USE</i>	Use of this product is only recommended together with the enclosed batch specific document which contains analytical data and stability information.
<i>PRODUCT DETERIORATION INDICATIONS</i>	The apparent features of the mixed control are similar to that of the blood. If it shows very dark red supernatant or unacceptable test results, it may indicate deterioration. Do not use the product if deterioration is suspected.

This lot was assayed by the test method described in the manufacturer's package insert.
These results are applicable to this lot and subsequent sublots, for which a capital letter has been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

This lot meets our specifications and is released for sale	APPROVED	Signed Date of Manufacture: Alicia Abellan 28 December 2018 Quality Control Dept.
	(APPROVED / NOT APPROVED)	

