

CERTIFICATE OF ANALYSIS	us-CRP TURBIDIMETRIC CONTROL
--------------------------------	-------------------------------------

REF	3125005	1 x 50 mL	3125080 (R1)	3925005 (CAL)	3925010 (CTROL) 1x2 mL	X
	3125049	1 x 25 mL	3125081 (R2)	3925020 (CAL)	3925120 (CTROL) 30x2 mL	X
	CT32500	1 x 50 mL		CT39250 (CAL)	CT39255 (CTROL) 2x2 mL	X

Cont **us-CRP Control**

LOT 39430
 2-8°C

 2018-03
IVD

PHYSICAL AND CHEMICAL CHARACTERISTICS

REAGENT	TEST	SPECIFICATIONS	RESULTS
us-CRP Control	APPEARANCE	Liquid. Clear and transparent.	ACCEPTED
	pH	8.2 ± 0.1	ACCEPTED

(ACCEPTED / N/A)

FUNCTIONALITY

REAGENT	TEST	SPECIFICATIONS	RESULTS
us-CRP Control	ASSIGNED VALUE	3,8 mg/L	ACCEPTED
	RANGE	3,04 - 4,56 mg/L	

(ACCEPTED / N/A)

STABILITY TEST

REAGENT	Days at 37°C	TOLERANCE	RESULT
us-CRP Control	10	After a stability study, the specifications remain unchanged	ACCEPTED

(ACCEPTED / N/A)

Components from human origin have been tested and found negative for the presence of antibody to HIV (1/2) as well as HBsAg and HCV. However, handle cautiously as potentially infectious.

This lot was tested by test methods described in the manufacturer's package insert.
 These results are applicable to this lot and sequential sublots, for which a capital letter have been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

This lot meets our specifications and is released for sale	APPROVED	Signed Date of Manufacture Marta Vilaró 2016-04-04 Quality Control Dept.
	(APPROVED / NOT APPROVED)	

