

<b>CERTIFICATE OF ANALYSIS</b>	<b>T3 EIA 96 Test</b>
--------------------------------	-----------------------

REF 

6107205	96 Test	x
---------	---------	---

<b>Cont</b>	<b>T3 Ab Coated Wells</b>	1
	<b>Standard Liquid</b>	1
	<b>Conjugate Concentrate</b>	1.3 ml
	<b>Conjugate Diluent</b>	13 ml
	<b>Antibody Reagent</b>	7 ml
	<b>TMB</b>	11 ml
	<b>Stop Solution</b>	11 ml

**LOT** RN-57193 2018-11  
 2-8°C **IVD**

**STANDARD CURVE**

Concentration (ng/mL)	O.D 450 nm
0.0	2.706
0.75	2.233
1.5	1.914
3.0	1.195
6.0	0.667
10.0	0.433

This standard Curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve

**CONTROL VALUES**

Control No.	Mean (ng/mL)	Range
1	1.00	0.423-1.30
2	2.00	1.31-2.55
3	2.89	2.10-3.51

This lot was tested by test methods described in the manufacturer's package insert.

These results are applicable to this lot and sequential sublots, for which a capital letter have been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

<b>This lot meets our specifications and is released for sale</b>	<b>APPROVED</b>	<b>Signed</b> <b>Date of Manufacture,</b> Alícia Abellan Quality Control Dept. 15/06/2017
	(APPROVED / NOT APPROVED)	

