


|                                |                      |
|--------------------------------|----------------------|
| <b>CERTIFICATE OF ANALYSIS</b> | <b>Linear URS-10</b> |
|--------------------------------|----------------------|

|            |         |                 |
|------------|---------|-----------------|
| <b>REF</b> | 7101005 | 100 Test strips |
|------------|---------|-----------------|

|   |           |
|---|-----------|
| <b>LOT</b>  | 9931 6576 |
|  | 2-30°C    |

|   |         |
|---|---------|
|  | 2019-08 |
| <b>IVD</b>  |         |

**FUNCTIONAL TEST WITH TEST URINE**

| PARAMETER                 | UNIT            | CONCENTRATION |           |            |           |           |             |       | RESULTS |
|---------------------------|-----------------|---------------|-----------|------------|-----------|-----------|-------------|-------|---------|
|                           |                 | Neg.          | +         | ++         | +++       |           |             |       |         |
| <b>Bilirubin</b>          |                 | Neg.          | +         | ++         | +++       |           |             |       | PASS    |
| <b>Urobilinogen</b>       | mg/dL<br>μmol/l | Norm.         | 2<br>35   | 4<br>70    | 8<br>140  | 12<br>200 |             |       | PASS    |
| <b>Ketones</b>            |                 | Neg.          | trace     | +          | ++        | +++       |             |       | PASS    |
| <b>Glucose</b>            | mg/dL<br>Mmol/L | Norm.         | 50<br>2,8 | 100<br>5,6 | 250<br>14 | 500<br>28 | ≥1000<br>56 |       | PASS    |
| <b>Protein</b>            | mg/dL           | Neg.          | trace     | 30         | 100       | 500       |             |       | PASS    |
| <b>Blood (lysed Erys)</b> | Erys/μl         | Neg.          | +         | ++         | +++       |           |             |       | PASS    |
|                           |                 |               | 5-10      | 50         | 300       |           |             |       |         |
| <b>Blood (intac Erys)</b> | Erys/μl         | Neg.          | +         | ++         | +++       |           |             |       | PASS    |
|                           |                 |               | 5-10      | 50         | 300       |           |             |       |         |
| <b>Nitrite</b>            |                 | Neg.          | Pos.      | Pos.       |           |           |             |       | PASS    |
| <b>pH</b>                 |                 | 5             | 6         | 6.5        | 7         | 8         | 9           |       | PASS    |
| <b>Specific Gravity</b>   |                 | 1,000         | 1,005     | 1,010      | 1,015     | 1,020     | 1,025       | 1,030 | PASS    |
| <b>Leukocytes</b>         | Leuk/μl         | Norm.         | 25        | 75         | 500       |           |             |       | PASS    |

(PASS / FAILURE)

This lot was tested by test methods described in the manufacturer's package insert.

These results are applicable to this lot and sequential sublots, for which a capital letter have been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

|   |                           |   |
|---|---------------------------|---|
| <b>This lot meets our specifications and is released for sale</b> | <b>APPROVED</b>           | <b>Signed -<br/>Date of Manufacture</b><br>Alicia Abellan<br>Quality Control Department<br>30 August 2017 |
|   | (APPROVED / NOT APPROVED) |   |

