



<b>CERTIFICATE OF ANALYSIS</b>	<b>Systemic solution</b> SOLUCION LIMPIADORA GENERAL GENERAL CLEANING SOLUTION SOLUÇÃO DE LIMPEZA GERAL
--------------------------------	--

<b>REF</b>	<b>KR18300</b>	10 x 50 mL	X
------------	----------------	------------	---

<b>Cont</b>	<b>LOT</b>	15668		2020-10	<b>IVD</b>	 2-25°C
-------------	------------	-------	---	---------	------------	--

### PHYSICAL AND CHEMICAL CHARACTERISTICS

TEST	SPECIFICATIONS	RESULT	
		ACCEPTED	X
Appearance	Liquid. Clear. Free of particles.	ACCEPTED	X
pH	$7 \pm 0.1$	ACCEPTED	X

(ACCEPTED / N/A)

This lot was assayed by the test method described in the manufacturer's package insert.

These results are applicable to this lot and subsequent sublots, for which a capital letter has been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

<b>This lot meets our specifications and is released for sale</b>	<b>APPROVED</b>	<b>Signed</b> <b>Date of Manufacture</b> Mayte Lázaro 03 Febrero 2017 Quality Control Dept.
---	-----------------	---

