


CERTIFICATE OF ANALYSIS	LIDA 500 CLEANING SOLUTION SOLUCION LIMPIADORA PARA ANALIZADORES LIDA 500 CLEANING SOLUTION FOR LIDA 500 ANALYZERS
--------------------------------	---

REF	1185010	1 x 1000 mL	X
------------	---------	-------------	---

Cont	LOT	15613		2020-07	IVD	 2-25°C
-------------	------------	-------	---	---------	------------	--

PHYSICAL AND CHEMICAL CHARACTERISTICS

TEST	SPECIFICATIONS	RESULT	
		ACCEPTED	X
Appearance	Liquid. Clear. Free of particles.	ACCEPTED	X
pH	≥ 9	13.17	X
Cross contamination rate	Diff% mean: ≤5%	1.33%	x

(ACCEPTED / N/A)

This lot was assayed by the test method described in the manufacturer's package insert.

These results are applicable to this lot and subsequent sublots, for which a capital letter has been assigned after the lot number.

This Certificate of Analysis is a computer printout and has therefore not been signed by hand.

This lot meets our specifications and is released for sale	APPROVED	Signed Date of Manufacture Mayte Lázaro Quality Control Dept. 13 July 2017
---	-----------------	--

